

जा.क्र.शा.आमव.रुज/भांडार/COVID-DPDC खरेदी/ १०० /२०२० दिनांक - 10 JUN 2021

दरपत्रक सुचना

प्रति,

विषय :- यंत्रसामुग्री/उपकरणे/ साहित्य खरेदीकरणेकरिता दरपत्रके सादर करणेबाबत

मा. जिल्हाधिकारी, जिल्हा आपत्ती व्यवस्थापन समिती, जळगांव यांचेकडून कोवीड-१९ अंतर्गत तसेच DPDC फंडामधून प्राप्त झालेल्या अनुदानातून या महाविद्यालय व संलग्नित रुग्णालयाकरिता सोबतच्या यादीतील नमुद तपशिलानुसार यंत्रसामुग्री/उपकरणे/ साहित्य खरेदीसाठी विहित वेळापत्रकानुसार संबंधित पुरवठादाकडून दरपत्रके मागविण्यात येत आहे.

तरी इच्छुक व पात्र पुरवठादारांनी त्यांची कमीत कमी दरांची दरपत्रके सिलबंद पाकिटांमध्ये अधिष्ठाता, शासकीय आयुर्वेद महाविद्यालय व रुग्णालय जळगांव यांचे कार्यालयात पोस्टाने अथवा समक्ष खालील नमूद अटी व शर्तीच्या अधिन राहून सादर करण्यात यावीत.

क्र.	बाब	कालावधी
1	दरपत्रक सादर करण्याचा कालावधी	दि ११.०६.२०२१ ते दि १९.०६.२०२१ सुटीचे दिवस वगळून कार्यालयीन कामकाजाच्या दिवशी सकाळी १०.०० ते सायंकाळी ५.०० वाजेपर्यंत.
2	दरपत्रक उघडण्याचा दिनांक व वेळ	दिनांक २२.०६.२०२१ रोजी दुपारी २.०० वा..
3	दरपत्रक उघडण्याचे ठिकाण	अधिष्ठाता, शासकीय आयुर्वेद महाविद्यालय व रुग्णालय, द्वारा श्री. गुलाबराव देवकर अभियांत्रिकी महाविद्यालय परिसर, सिरसोली रोड, जळगांव. ४२५००१.

अटी व शर्ती

- दिनांक १९-०६-२०२१ रोजी सांय ५.०० वाजल्यानंतर प्राप्त होणा-या दरपत्रकांचा विचार केला जाणार नाही.
- दरपत्रकाच्या पाकीटावर यंत्रसामुग्री/उपकरणे/ साहित्य खरेदी (भाग १ अथवा भाग २) दरपत्रक सादर करणेबाबत असा स्पष्ट उल्लेख करणे आवश्यक आहे.
- दरपत्रकात कोणत्याही प्रकारची खाडाखोड किंवा ओव्हररायटींग असता कामा नये.
- पुरवठादाराने GST ची नोंदणी करणे आवश्यक राहिल. तसेच दरपत्रके संस्थेची नोंदणी क्रमांक असलेल्या लेटर हेडवर दरपत्रक सुचनेनुसार विहित नमुन्यात सादर करणे आवश्यक राहिल.
- दरपत्रक सादर करणा-या दरपत्रकधारकाचा किंवा संस्थेचा दरपत्रक प्रक्रियेत भाग घेणा-या अन्य संस्थेमध्ये थेटस्वारस्य असू नये, तसे आढळल्यास दोन्ही किंवा अधिक दरपत्रके नाकारण्यात येतील. याकरीता सोबत जोडण्यात आलेली हमीपत्र दरपत्रकधारकास सादर करणे बंधनकारक राहिल. (हमी पत्र विहित नमुन्यात सादर करावे)
- दरपत्रक धारकांनी केवळ एकच दरपत्रक सादर करावे. दरपत्रक सादर केलेल्या दरपत्रक धारकाचा दरपत्रक प्रक्रियेत भाग घेतलेल्या इतर संस्थासोबत प्रत्यक्ष/अप्रत्यक्ष संबंध वा सहभाग नसावा. तसे आढळल्यास दरपत्रक नाकारण्यात येतील.
- दरपत्रकासोबत GST, Pan Card, Shop Act Licence, renewal etc. प्रमाणपत्रे सादर करणे आवश्यक राहिल.

८. देयकाच्या रकमेतुन शासन नियमातील तरतुदीनुसार GST/ Income Tax कपात करण्यात येईल. सदर टॅक्स मधुन सुट अनुज्ञेय असल्यास त्याबाबतचा पुरावा सादर करणे आवश्यक राहील.
९. दरपत्रके उघडण्याच्या वेळी कंपनीच्या एका प्रतिनिधीस स्वखर्चाने हजर राहण्याची मुभा राहील.
१०. दरपत्रकानुसार मंजूर करण्यात आलेल्या दराव्यतिरिक्त इतर कोणतेही दर देय होणार नाहीत.
११. पुरवठा करण्यात येणारी यंत्रसामुग्री/उपकरणे/ साहित्य संस्थेस पोहोच करण्याची जबाबदारी पुरवठादाराची राहील. त्यासाठी कोणतेही अतिरिक्त शुल्क अदा केले जाणार नाही.
१२. शा.निर्णन उद्योग व उर्जा विभाग दि.०१.१२.२०१६ मध्ये नमूद केल्यानुसार पुरवठादाराच्या बाबतीत खरेदीच्या कोणत्याही टप्प्यावर अनियमितता /गैरप्रकार/नैतिक अधःपतन इतरही बाबी निदर्शनास आल्यास त्याबाबत शासन निर्णयामध्ये नमूद केल्यानुसार कार्यवाही करण्यात येईल.
१३. यशस्वी दरपत्रक धारकास पुरवठा आदेश निर्गमित झाल्यानंतर यंत्रसामुग्री/उपकरणे/ साहित्याचा पुरवठा करणे आवश्यक राहील. तसेच पुरवठा केल्यानंतर ३० दिवसाचे आत सदर यंत्रसामुग्री/उपकरणे/ साहित्य खराब असल्याचे निदर्शनास असल्यास सदर यंत्रसामुग्री/उपकरणे/ साहित्य स्वखर्चाने परत नेऊन त्याऐवजी नवीन यंत्रसामुग्री/उपकरणे/ साहित्याचा पुरवठा ३० दिवसाचे आत करावा लागेल.
१४. यशस्वी दरपत्रकधारकास त्यांचे देयकांची रक्कम ऑनलाईन पध्दतीने त्यांचे बँक खात्यात जमा करण्यासाठी विहित नमुन्यातील राष्ट्रीयकृत बँक खाते प्रमाणीकरण प्रपत्र, पॅनकार्ड झेरोक्स प्रत व रद्द केलेला धनादेश इ. देणे आवश्यक राहील. त्याशिवाय देयकाची रक्कम अदा करणे शक्य होणार नाही.
१५. अधिष्ठाता, शासकीय आयुर्वेद महाविद्यालय, जळगांव हे सदर दर पत्रक स्विकारणे, रद्दकरणे, यशस्वी दरपत्रक स्विकारणाच्या पूर्वी किंवा यशस्वी दरपत्रक स्विकारण्यात आल्यानंतर देखिल कोणत्याही वेळी रद्द करण्याचे अधिकार तसेच खरेदी प्रक्रियेच्या अटी व शर्तीमध्ये कोणताही बदल करण्याचा अधिकार राखुन ठेवत आहे.
१६. पुरवठादाराने सदर खरेदी संदर्भातील आवश्यक माहिती यासोबतच्या प्रपत्रात (Annexure-B) मध्ये सादर करावी



(वैद्य मिलींद भा. निकुंभ)

अधिष्ठाता,

शासकीय आयुर्वेद महाविद्यालय व रुग्णालय,
जळगांव.

उपरोक्त दरपत्रकाची सुचना महाविद्यालयाची वेबसाईट www.gacjalgaon.in वर प्रसिध्द करण्यात आली आहे.

अर्जासोबत खालील कागदपत्रे जोडण्यात यावीत.

१. विहित नमुन्यातील स्वाक्षरीत हमीपत्र
२. आधारकार्ड झेरोक्स स्वसाक्षात्कृत प्रत
३. पॅनकार्ड झेरोक्स स्वसाक्षात्कृत प्रत
४. शॉपअॅक्ट प्रमाणपत्र झेरोक्स स्वसाक्षात्कृत प्रत.
५. GST प्रमाणपत्र झेरोक्स स्वसाक्षात्कृत प्रत.
६. व्यवसाय नोंदणी प्रमाणपत्र छायांकित स्वसाक्षात्कृत प्रत.
७. संस्थेचा रद्द केलेला धनादेशाची झेरोक्स स्वसाक्षात्कृत प्रत

Annexure - B

कंपनीच्या लेटर हेडवर खालील नमुद तपशिलाप्रमाणे माहिती भरणे.

प्रति,
अधिष्ठाता,
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय,
जळगांव.

क्र.	तपशिल	आवश्यक माहिती व प्रमाणपत्रे जोडण्यात आलेली आहेत किंवा कसे
१	पुरवठादाराचे नांव	
२	पुरवठादाराचा पत्ता, ई-मेल, दुरध्वनी इ.	
३	पुरवठादाराचा नोंदणी क्रमांक (Registration No.)	
४	GST No.	
५	PAN No.	
६	सदर वस्तुंचा इतर शासकीय अथवा निमशासकीय कार्यालयास पुरवठा केला असल्यास अशा कार्यालयाचे अथवा पुरवठा आदेशांची छायांकित प्रत, पत्ता व दुरध्वनी क्रमांक नोंदविण्यात यावा.	
७	हमी प्रत	
८	इतर अटी व शर्ती असल्यास	

स्थळ -

दिनांक -

पुरवठादाराची अधिकृत स्वाक्षरी शिक्कांसहित

हमीपत्र - १ (Undertaking)

प्रति,
अधिष्ठाता,
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय,
जळगांव.

मी (संस्था चालकाचे व संस्थेचे नांव)

पत्ता

लेखी हमी देतो की, आपल्या कार्यालयामार्फतखरेदी करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी भाग घेत असून माझे किंवा माझ्या संस्थेचे खरेदी प्राधिकारी/ खरेदी समिती वा कार्यालयाशी कोणत्याही प्रकारचे हितसंबंध/ संघर्ष नाहीत. तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ -

दिनांक -

पुरवठादाराची अधिकृत स्वाक्षरी शिक्कांसहित

हमीपत्र - २ (Undertaking)

प्रति,
अधिष्ठाता,
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय,
जळगांव.

मी

पत्ता

लेखी हमी देतो की, आपल्या कार्यालयामार्फतखरेदी करणेसाठी राबविण्यात येत असलेल्या दरपत्रक प्रक्रियेत मी/ माझी संस्था केवळ एकच दरपत्रक सादर करीत आहे. त्याचप्रमाणे दरपत्रक भाग घेणा-या अन्य संस्थांसोबत माझा/माझ्या संस्थेचा कोणत्याही प्रकारचा कोणत्याही प्रकारचा प्रत्यक्ष/ अप्रत्यक्ष संबंध वा सहभाग नाही. तसे आढळून आल्यास माझे दरपत्रक नाकारण्यात यावे वा अपात्र समजण्यात यावे.

स्थळ -

दिनांक -

पुरवठादाराची अधिकृत स्वाक्षरी शिक्कांसहित

शासकीय आयुर्वेद महाविद्यालय व रुग्णालय जळगांव

श्री गुलाबराव देवकर अभियांत्रिकी महाविद्यालय परिसर, सिरसोली रोड, जळगाव - ४२५००२

ईमेल- gacjalgaon@gmail.com

यंत्रसामुग्री/उपकरणे/ साहित्याची यादी (भाग - ०१)

क्र.	यंत्रसामुग्री/उपकरणे/ साहित्याची नांव	स्पेसिफिकेशन	संख्या (Qty)
1	Bipap (Niv)	1- Pressure Range (cm H2O)- 4-25 CM H2O 2- Operating Mode- CPAP, BIPAP, PIP, HFNC 3- Brand- Yashka 4- Data Storage Capacity- Display-compliance VIC, 1-, 7- and 30 day averages. 5- Dimensions- 30 x 25 x 25 6- Weight- 20 kg (without compressor) 7- Filter- HME Filter 8- Starting Ramp Pressure- 4 to EPAP (Patient Adjustable) 9- Compliance meter- Breathing Detection 10- Electrical requirements- 100-240 Vac, 50/60 Hz	10
2	PPE Kit		5000
3	Mask	Tripple layer	10000
4	Mask	N 95	5000
5	Sanitizer		500 Ltr
6	Floor cleaner		500 Ltr
7	Laboratory Refridgerator	Single/double freezer, defrost time : frost free, inverter compressor, capacity 400 liter, no.of door-2, star-3, coolpad, stabilizer built in. Temp -18 to 8	01
8	Daikin Air Conditioner 2 tone capacity	2 Tonnes with copper piping, Remote control, inverter Hot & Cold condensor 24 VH. Ezo Split AC.	02
9	Centrifuge Machine	CENTRIFUGE CLINICAL FOR 12 TUBESS Technical specifications Tube capacity : no 12 or more :size 5-15 ml Should have a digital timer, body should be made of strong fabricated & corrosion resistant steel. Control panel- for start/stop switch & speed indicator with timer and protective fuses. Door interlock, Maintenance- free brushless drive motor with exact speed preselection and display . speed range 100 to 6000 rpm and above ,accuracy 1 rpm .RPM: up to 4000-6000 Power supply Power input to be 220 -240 VAC.50	02

10	Semi Automatic Bio Chemistry Analyser	Semi automatic bio chemistry EC5X with incubator and warranty 3 yrs	01
11	AUTOMATED HEMATOLOGY ANALYSER (3 PART)	automatic and manual Hematology Analyser (3 PART)	01
12	PT/INR D-Dimer Test machine	-	01
13	Laboratory Reagents	As per list	-
14	O2 Controlling Pannel	- Liquid Oxygen - Jumbo Oxygen Cylinder controller & Shifter Atomic	01
15	Bed Sheets	Single Bed 7' x 4' Green/ Blue color	500
16	Solapuri Chadri	Single 7' x 4'	250
17	Blankets	Single 7' x 4'	250
18	Dry clean / Laundry Machine	Washer Steamer	50 Ltr
		Dryer	50 Ltr
19	Binocular Research Microscope	Microscope Type- pathological binocular non hinged type with built in light and with light intensity regulator Eye Piece with magnification- monocular 5 x Objective Magnification- 10 x Numerical Apperture of Objective- 0.5 Type of lamp for Illumination- led Size of Stage- 120x140 Coarse and Fine Movement of stage, Co-axial focussing ,	

(वैद्य मिलीद भा. निकुंभ)

अधिष्ठाता,

शासकीय आयुर्वेद महाविद्यालय व रुग्णालय,
जळगांव.

महाराष्ट्र शासन
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय जळगांव
श्री गुलाबराव देवकर अभियांत्रिकी महाविद्यालय परिसर, सिरसोली रोड, जळगाव - ४२५००२
ईमेल - gacjalgaon@gmail.com

Laboratory Reagents List

अ.क्र.	रसायने / वस्तुचे नांव	Qty
1	Micro Pipette 1 to 10	1
2	Micro Pipette 10 to 100	1
3	Micro Pipette 100 to 500	1
4	Uristix	20 Box
5	E.D.T.A. Bulbs	1 Box
6	Sr. Creatinine Test kit	75 ml x 06 Kits
7	RA Test kit	20 Test x 1 kits
8	Cotton bundles	500 gms x 1nos.
9	HCG/ UPT Test Kit	50 Kits
10	Sr. Bilirubin Test Kit	75 ml x 1 Kits
11	SGOT Test kit	75 ml x 1 Kits
12	SGPT Test Kit	75 ml x 1 Kits
13	Sr. Urea Test Kit	75 ml x 1 Kits
14	Sr. Uric Acid Test Kit	25 ml x 1 Kits
15	Sr. Cholesterol kit	75 ml x 1 Kits
16	Sr. Triglycerides Kit	75 ml x 1 Kits
17	Sr. Albumin kit	150 ml x 1 Kits
18	Sr. Total Protein kit	150 ml x 1 Kits
19	Glucose Reagent	01 litre x 01 Nos.
20	Celltac Chemi HbA1C Measuring kit	50 Test x 01 Box
21	ERBA CRP Kit	50 Test x 01 Box
22	Latex Examination Gloves	5 Box
23	Disposable Syringes (2.5 ml)	100 Syringes x 2 Box
24	Disposable Syringes (5 ml)	100 Syringes x 2 Box
25	Glass Test Tubes	100 Test Tubes x 1 Box
26	WIDAL Test Kit	05 ml x 1 Kit
27	Coverslip Box	1 Box
28	Leishmans Stain	250 ml x 1 Nos.
29	Deionised / Distilled Water	1 liter x 1 Nos.
30	VDRL / RPR Test Kit	10 Test
31	ASO Test Kit	01 Box
32	Differential Ph Paper	01 Box


Dr. Milind B. Nikumbh
Dean, Govt. Ayurved College
Jalgaon

महाराष्ट्र शासन
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय जळगांव

श्री गुलाबराव देवकर अभियांत्रिकी महाविद्यालय परिसर, सिरसोली रोड, जळगाव - ४२५००२
ईमेल - gacjalgaon@gmail.com

यंत्रसामुग्री/उपकरणे/साहित्याची यादी (भाग - २)

क्र	यंत्रसामुग्री व साधनसामुग्रीचे नांव	Specifications	नगसंख्या
1	Multi- Para monitor	12.1" color TFT LCD screen with Multi channel waveforms display Rechargeble high energy built in Li Battery easy to learn easy to use simplified menus delicated function keys Standard Parameters ECG Lead, SpO2, Resp, 2-Temp, HR	5
2	ECG Machine (12 lead)	Microprocessor controlled 12 channel ECG Machine with interpretation. FEATURES:- *Compact 12 channel with Interpretation. Parameters Display on LCD. Compact, lightweight, portable, Easy to use & service. High Fidelity Recording. Maintenance free digital thermal array printer hence free from friction, non linearity damping stylus heating problems Clear ECG tracings & easy paper loading. Electronically controlled feather touch key operations. In built rechargeable battery with charger, Automatic & Manual modes. TECHNICAL SPECIFICATIONS, Leads : Standard 12 Leads Input Impedance : 50M ohm Time Constant : more than 3.2 seconds Noise : +1/4 LSB (less than 1 micro volt peak to peak) Filters : Digital AC 50 Hz Writing Device : Electronic thermal array Printer Recording Paper : Thermal 50 MM. 30 mtrs	2
3	Multi-Para monitor with NIBP	12.1" color TFT LCD screen with Multi channel waveforms display Rechargeble high energy built in Li Battery easy to learn easy to use simplified menus delicated function keys Standard Parameters ECG Lead, SpO2, Resp, 2-Temp, HR, NIBP	10

4	ICU BED With Mattress	2-Section Tuck Away ABS Side rails with Safety Lock 3-Section Perforated CRC sheet and backrest having X-Ray translucent top with cassette tray. Removal and Lockout ABS Head & Foot Board Central Brake System, Mounted on 125mm Dia castor. ABS Corner Bumper Drainage Bag Hooks Four I.V. Pole Sockets I.V. Pole Storage Finish: Epoxy Powder Coated Available with removable I.V. Pole	10
5	IV Stand	IV Saline Stand, Size (Dimensions in Length X Breadth X Height): 50 X 50 X 200 Cm Number of Wheel- 3 Castor size- 5 cm Material of stand- Stainless steel	50
6	Auto refracto keratometer	Objective Refractometer Mode- KR-800 / RM-800 / Sphere Range- -25D to +22D (0.12D/0.25D step) Cylinder Range- 0D to +10D (0.12D/0.25D step) Axis Range- 0° to 180° (in 1° or 5° steps) Minimum Measurable pupil Diameter- ϕ 2mm Corneal curvature radius- 5.00 to 10.00mm (0.01mm step) Corneal refractive power- 67.50D to 33.75D (0.12D/0.25D steps) (where, corneal refractive power =1.3375) Corneal astigmatic refractive power- 0D to $\pm$ 10D (0.12D/0.25 D steps) Corneal astigmatic axial angle- 0° to 180° (1°/ 5°steps) 8.5-inch LCD touchscreen panel	1
7	Ophthalmic microscope (appasamy)	Binocular tube- Parallel 45° inclined binocular tube Objective lens- F=200mm Eyepieces- 12.5x Magnification change type- Drum type with 5 changes Total magnification- 3.4x, 5.3x, 8.8x, 13.6x, 21.2x Field of view- ϕ 66.3, ϕ 42.3, ϕ 26.3, ϕ 17.1, ϕ 10.3mm X-Y Translator (option)- Range: $\pm$ 25mm	1

		Length- 1st arm:400mm; 2nd arm:560mm Rotational angle- 1st arm:360 °; 2nd arm:300 ° Base dimension- φ 740mm (30") Primary voltage- AC100-120V or AC220-240V Power frequency- 50/60Hz Maximum power consumption- 280VA Weight- 97kg	
8	Labour Table (Telescopic and side tilting)	Telescopic labour table full ss, side Tilting Facility, Head lower facility with mattress, 1800mm x 675 mm x 6-800mm approx, wheels -04, foldable-yes	2
9	Instrument cabinet(glass)	60 Inch x 48 Inch 18 inch (SS) glass door	5
10	Bed side lockers	Having one Box and one drawer S. S. Top with three sides raised edges Dimension (LxWxH)- 16 x 14 x 30 Inch Material Finish- Epoxy Powder Coated	125
11	Infusion Pump	Pumping Mechanism, Curvilinear Peristaltic, Compatible with IV sets of any standard, Flowrate- 0.1-1500ml/G (in 0.1ml/h increments) Bolus volume 1-20 ml, Accuracy- +- 3%, ml/h, drop, kin, time-based, Battery 5-6 hr at 30 ml/h	5
12	Syringe Pump	No of syringe 2.0, 6.0, 4.0 with 3 kg, Mode Drug delivery, Type Continuous flow front loading syringe	5
13	Examination couch	6' x 2 ½' with mercury BP app. Tray, Drawer & shelf with cushion with foot step	7
14	Auto-clave (Double drum)	14" x 9" - 6 drum Capacity	2
15	Surgical cautery (under water cutting)	Bipolar, 400 wat, Underwater cutting	1

16	NST (Fetal Toco-graph Machine)	Display- 7-10 inch, tilt-90 degree upto parameters, FHR, Toco, event marker automatic fetal movement detection, Automatic CTG reporting Scoring facility interface, twin monitoring facility as std feature, 100 - 240 v Built in pc network can be connected to control monitoring system	1
17	Photo therapy unit (top and under surface)	Photo Therapy unit shall be equipped with white light for observation purpose Power consumption in watts- upto 45 Number of castors- 4 Double surface unit with bassinet (made of glass) Number of brakes- 2 Natural cooling	5
18	Neonatal Warmer & incubator unit	Overall height of equipment – 800 millimeter Heat source configuration- 60 degree angle adjustment must be possible. Heat dissipation- should maintain desired set temperature and uniform heat distribution in baby cradle. Overall Length of equipment – 700 millimeter Protection- OVP, earth leakage protection. Thermal reflector to fix the skin probe- 50 units to be supplied with each equipment unit. Power consumption- 400 watt.	5
19	Linen Trolley with canvas bag	• Superior Quality Canvas Bag. • Framework- SS Tubular Framework (304 Grade) • Castor- 125 mm dia without brake • No. of Bows- 910mm(H)x510mm(W) • No. of Legs- Three Legs • Wheel Diameter- 100 mm • Color- SS Finish	5
20	500/300 MA X-ray machine digital	Model - 22 KW, Application- Radiographic, Generator Type- High Frequency Output (Maximum 400 kv), Rotor Supply- Low Speed Starter (Optional Dual Speed Starter), Tube Calibration- Standard Feature, Image Receptors- Up to 5, Tomography Mode- Standard Feature, Auxiliary Collimator Power Supply- Standard Feature	1

21	USG (3D/4D) Machine	Screen features an innovative 21.5" full HD high resolution, touch monitor and multi-touch based User Interface with interactive dynamic software menu., MODES: B-Mode, M-Mode, PW Doppler, Color Flow Doppler, HD-Flow, Power Doppler, AUTO IMAGE OPTIMIZATION FEATURES: Automatic Optimization B-Mode and PW Doppler, Auto TGC, Auto PRF, Flow Profiles all color and Doppler modes, IMAGING TECHNOLOGIES: Coded Harmonic Imaging with Pulse Inversion Technology, SRI (Speckle Reduction Imaging), CrossXBeamCRI™ (Compound Resolution Imaging), Histogram analysis, AUTOMATED CALCULATION FEATURES: SonoBiometry, SonoFHR, Real-time Doppler calculations, IMAGING FORMATS: Steer, Virtual Convex, Trapezoid, Wide Sector, BetaView, HD Zoom & Pan Zoom, MEASUREMENT, CALCULATIONS AND WORKSHEET/REPORTS available across a full range of clinical applications, SPECIALIZED REPORTING PACKAGES: Multi-gestational Calculations, IOTA (International Ovarian Tumor Analysis): LR2, Simple Rules and ADNEX Model (not available in all countries), IDEA (International Deep Endometriosis Analysis) Protocol & Scan Assistant Guideline, STORAGE AND CONNECTIVITY: 1 TB Hard Drive, DICOM 3.0 Connectivity	1
22	Panchakarma Instrument set	<i>As per list</i>	2
23	Tablet making machine	GMP Model Material: S S 304/316, 24 stations double rotatory, Extra set of punches of 8 mm, 10 mm and 12 mm , Fully automatic	1
24	Disintegrator	Material: S S 304/316 all contact part and hammer 7.5 hp motor/ 3 phase Extra pair of sieves GMP Model	1
25	Tray Dryer	S.S. 304/316 Material 24 S S tray capacity (S S Tray 16" X32" X 1 1/4") Digital Temperature Control Unit Maximum temperature 150 O C GMP Model	1

26	Vibro shifter	All contact parts must be of S S304/31 Noise Free High Speed Capacity (800-900 kg/hr) Single Decker 30 inch in diameter GMP Model Extra pair of sieves of 60 and 80 mesh size	1
27	Mass Mixer	Material : S S 304/316 Working Capacity : 100 kg 2 hp motor / 3 phase S S dust proof cover Titling facility G M P Model	1
28	Liquid Filling Machine	Semi automatic GMP model All contact part S.S. 304/S.S. 316. Filling range from 50, 100 and 200 ml. (Twin head) Motor : 0.5 hp	1

	 (वैद्य मिलींद भा. निकुंभ) अधिष्ठाता, शासकीय आयुर्वेद महाविद्यालय, जळगांव.
--	---

महाराष्ट्र शासन
शासकीय आयुर्वेद महाविद्यालय व रुग्णालय जळगांव
श्री गुलाबराव देवकर अभियांत्रिकी महाविद्यालय परिसर, सिरसोली रोड, जळगाव - ४२५००२
ईमेल- gacjalgaon@gmail.com

Panchakarma Instrument Set

Sr.No.	Instruments Name	Instruments Details (Specification)	Qty
1	Massage Droni	Massage Droni- Top with single piece/single joint, dedical wood with standard carved wooden stand, export superior quality (size-8.5' length x 30' width x 3.5 thickness Single piece, Single joint.	1
2	Shirodhara stand	Shirodhara stand (Carved Wood) Vessel Height adjustable without vessel (185 2 cms height)	1
3	Semi Automatic Boiler	Semi Automatic Boiler- Autoclave Stainless steel insulated Standard Quality model- Electric 200 wts, 10 ltr capacity	1
4	Massage Droni	Massage Droni- (Pixhichil oil bath, Dhara, Abyanga) Fiberglass material make heavy duty and attractive Wooden finishing Export superior quality Size 8.5' x 30' x 3.5'	1
5	Bronze metal vessel	Bronze metal kindi (light duty aprox 800 gms) with handle 500 ml capacity	1
6	Shirovasti Cap	Shirovasti Cap- Adustable leather cap made out of calf leather. Size length 80 cm x width 20 cm	1
7	Bronze Metal Vessel	Bronze metal vessel (light duty) 1.2 kgs (uruli shape with adjustable valve) 2.5 lts capacity	1
8	Nasya Yanthram	Nasya Yanthram (Bronze) gokarnam made up of white bronze. Diamter 4cm weitht 44 gms	1
9	Steam Chamber	Lying type Steam Chamber Heiht 43' (108 cms) Length 94' (240 cms) Width 31' (78 cms) Wood Teak wood	1
10	Spinal Bath FRP	Spinal Bath FRP Length 45' (115 cms) Width 24' (60 cms) Heiht 10' (25 cms) Capacity 40 ltr Sturd FRP, Wood Plin rose wood or jungle wood finish	1
11	Wooden foot step	Wooden foot step for export quality double step	1
12	Matra Basti	Matra Basti syring steel (100 ml) these have stainless steel body with head.	1